

Optical calibration system of NIPR for aurora and airglow observations

*小川 泰信^{1,2,3}、門倉 昭^{1,2,3}、江尻 省^{1,2}

*Yasunobu Ogawa^{1,2,3}, Akira Kadokura^{1,2,3}, Mitsumu K. Ejiri^{1,2}

1. 国立極地研究所、2. 総合研究大学院大学、3. 情報・システム研究機構 データサイエンス共同利用基盤施設

1. National Institute of Polar Research, 2. The Graduate University for Advanced Studies, SOKENDAI, 3. Joint Support-Center for Data Science Research, Research Organization of Information and Systems

Calibration of optical instruments is important for accurate measurement of the absolute emission intensity of aurora and airglow. The National Institute of Polar Research (NIPR), Japan, has been operating a facility for the optical calibration which consists of three independent calibration systems to share with collaborating researchers. This paper introduces an outline of the facility and specifications of each system, calibration procedures, and examples of calibration results. With this facility, we can obtain the calibration data for the absolute sensitivity of optical instruments with an accuracy of about 2% within a wavelength range between 420 and 1050 nm. In addition, it is possible to calibrate transmission of optical filters and also relative sensitivity of optical instruments as a function of wavelengths. Current facility, upgraded in 2012, has been used by many researchers to calibrate their various optical instruments.

Reference:

Ogawa, Y., A. Kadokura, M. K. Ejiri, Optical calibration system of NIPR for aurora and airglow observations, Polar Science, doi:10.1016/j.polar.2020.100570, 2020.

キーワード : Aurora、Airglow、Calibration

Keywords: Aurora, Airglow, Calibration